

## 紡織實驗室

**Test Report** 報告號碼: TXB0507 /2020 /PL 日期: 2020 年 12 月 21 日 頁數: 二 之 一 頁

台灣京富醫材有限公司  
彰化縣花壇鄉花橋街 83 巷 18 號

我們依照客戶的要求, 根據申請者送交的樣品進行測試如下:

樣品敘述 : 熔噴布(黑)25gms/m<sup>2</sup>

顏色 : 黑色

成份 : PP

申請者 : 台灣京富醫材有限公司

收件日期 : 2020 年 11 月 06 日

測試週期 : 2020 年 11 月 06 日至 2020 年 12 月 21 日

測試項目&測試方法 : 依申請者要求進行下列測試

測試結果 : 請參照下一頁結果

台灣檢驗科技股份有限公司

張家豪

張家豪  
副理



"Unless otherwise stated the results shown in this test report refer only to the sample(s) tested, and such sample(s) are retained for 1 months only." This test report cannot be reproduced, except in full, without prior written permission of the Company.

"This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com.tw/Terms-and-Conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com.tw/Terms-and-Conditions>.

Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law."

紡織實驗室

Test Report 報告號碼: TXB0507 /2020 /PL 日期: 2020 年 12 月 21 日 頁數: 二 之 二 頁

測試結果:

細菌過濾效益

測試項目	測試結果	測試方法
細菌過濾效益 (BFE)(%)	1	>99.9
	2	>99.9
	3	>99.9
	4	>99.9
	5	99.9
		ASTM F2100-19 9.1 ASTM F2101-19

備註: 樣本數量: 5  
 測試面: 任意面  
 測試面積: 49 cm<sup>2</sup>  
 採樣流量: 28.3 Liter/min  
 陽性對照組的平均菌落數: 1960 CFU  
 陰性對照組的平均菌落數: 0 CFU  
 平均粒徑大小: 3.3 μm

樣品編號	Anderson 各階菌落數						總菌落數 (CFU)
	第 1 階 (7 μm)	第 2 階 (4.7 μm)	第 3 階 (3.3 μm)	第 4 階 (2.1 μm)	第 5 階 (1.1 μm)	第 6 階 (0.65 μm)	
01	0	0	0	0	0	0	0
02	0	0	0	0	0	0	0
03	0	0	0	0	0	0	0
04	0	0	0	0	0	0	0
05	0	0	0	0	0	2	2

ASTM F2100-19 醫療口罩等級要求

試驗項目	Level 1	Level 2	Level 3
細菌過濾效益(BFE) (%)	>= 95	>= 98	>= 98

測試由其他實驗室執行

\*\*\* 結 束 \*\*\*

"Unless otherwise stated the results shown in this test report refer only to the sample(s) tested, and such sample(s) are retained for 1 months only." This test report cannot be reproduced, except in full, without prior written permission of the Company.  
 "This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com.tw/Terms-and-Conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com.tw/Terms-and-Conditions>.  
 Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law."